



DECEMBER 11-12, 2025
۲۱ و ۲۰ آذر ماه ۱۴۰۴



دومین کنگره
پلازما پزشکی ایران

The 2nd Congress on Plasma Medicine

دبیرخانه دائمی کنگره
پلازما پزشکی ایران
www.plasmamedsym.ir



Simulation of a Multi-Hollow Surface Dielectric Barrier Discharge Device for Cold Plasma Generation

Parisa Azarinoor¹, Mahdiyeh Bakhtiyari Ramezani², Fatemeh Shakeri¹, and Pedram Nasiri*¹

¹Department of Plasma Physics, Plasma Technology Development Company, Tehran, Iran

²Plasma and Nuclear Fusion Research School, Nuclear Science and Technology Research Institute (NSTRI), Tehran, Iran

OPEN ACCESS

Citation:

Azarinoor P, Bakhtiyari Ramezani M, Shakeri F, Nasiri P. Simulation of a Multi-hollow Surface Dielectric Barrier Discharge Device for Cold Plasma Generation. *Iran Biomed J. Supplementary* (2-2026): 15.

ABSTRACT

Introduction: This research involved simulating a multi-hollow surface dielectric barrier discharge (MSDBD) device to examine the distribution of the electric field and its potential for plasma ignition. The aim was to identify essential areas where the electric field intensity increases, which play a crucial role in establishing conditions necessary for plasma formation at atmospheric pressure.

Materials and Methods: A two-dimensional model of the MSDBD device was developed using COMSOL Multiphysics 6.3. The geometry and material properties were configured based on experimental samples described in the scientific literature. In this model, a voltage of 5 kV was applied to the nickel and gold electrodes embedded within an alumina insulating layer. To accurately capture variations in the electric field, a highly refined mesh was employed, with element sizes ranging from 0.03 mm to 5.3 mm.

Results and Discussion: The simulation revealed an intense electric field buildup inside the cavities, particularly close to the walls. The peak electric field strength near the walls reached around 9 MV/m, whereas at the center, it dropped to roughly 5.8 MV/m. Both levels were significantly above air's breakdown limit at normal pressure –about 3 MV/m–suggesting that plasma could form in these regions. Changes in relative permittivity, electrode spacing, or surface shape noticeably influenced the field spread, highlighting the significance of material selection and exact sizing when building the MSDBD.

Conclusion: The simulations indicated that the designed MSDBD setup is capable of producing plasma when subjected to a voltage of 5 kV. This model serves as a valuable quantitative resource for predicting plasma dynamics, enhancing device design, and minimizing the need for continuous experimental trials. Such evaluations play a significant role in developing local technical expertise and advancing the biomedical uses of MSDBD-powered cold plasma sources.



This article is licensed under a Creative Commons Attribution-NonDerivatives 4.0 International License.

Keywords: COMSOL Multiphysics, Cold plasma, Electric field simulation, MSDBD

Corresponding Author: Pedram Nasiri

Department of Plasma Physics, Plasma Technology Development Company, Tehran, Iran; E-mail: Pedramn27@gmail.com



Iranian Biomedical Journal Supplementary (February 2026): 15