



DECEMBER 11-12, 2025  
۲۱ و ۲۰ آذر ماه ۱۴۰۴



دومین کنگره  
پلازما پزشکی ایران  
The 2<sup>nd</sup> Congress on Plasma Medicine

دبیرخانه دائمی کنگره  
پلازما پزشکی ایران  
www.plasmamedsym.ir



# Cold Plasma-Enhanced Nanostructures for Next-Generation Biomedical Plasmonic Sensors

Rahele Aboltaman

Department of Physics, Faculty of Science, Arak University, Arak, Iran

## OPEN ACCESS

### Citation:

Aboltaman R. Cold Plasma-Enhanced Nanostructures for Next-Generation Biomedical Plasmonic Sensors. *Iran Biomed J. Supplementary* (2-2026): 28.



This article is licensed under a Creative Commons Attribution-NonDerivatives 4.0 International License.

**Keywords:** Ag nanostructures, Argon atmospheric plasma jet, SERS plasmonic biosensor

## ABSTRACT

**Introduction:** Low-cost, portable surface-enhanced Raman scattering (SERS)-nanoplasmonic biosensors enable early, molecular-level diagnostics in precision medicine. Treatment with atmospheric-pressure plasma treatment enhances surface properties, allowing for highly sensitive, non-destructive Raman detection of biomolecules such as lysine at concentrations as low as  $10^{-14}$  M. The effects of plasma exposure parameters on sensor performance were also evaluated.

**Materials and Methods:** Silver nanostructure-based SERS plasmonic biosensors were fabricated on plasma-treated silicon via electroless deposition. The sensors were further modified by atmospheric-pressure plasma treatment (power: 35 W, output gas rate: 6 Lit/min, treatment time: 8 min).

**Results and Discussion:** At low concentrations, the Raman signal of lysine is often obscured by fluorescence. However, applying 3  $\mu$ L of  $10^{-4}$  M lysine solution to the nanoplasmonic SERS sensor allows detection even at this concentration, enabling identification of lysine, which is a biomolecule implicated associated with metabolic disorders.

**Conclusion:** Optimized argon plasma treatment of silver nanostructures enhances sensor performance by increasing surface energy, introducing nanoscale defects and roughness, and promoting denser nanodendritic growth that generates more active hotspots. Additionally, increased hydrophilicity facilitates better adsorption of target molecule, thereby improving the efficiency of plasmonic biosensors.

**Corresponding Author:** Rahele Aboltaman

Department of Physics, Faculty of Science, Arak University, Arak, Iran; E-mail: fatemeh.r-aboltaman@araku.ac.ir



**Iranian Biomedical Journal Supplementary (February 2026): 28**